



Smartway Electronics Pvt. Ltd.

SMARTWAY PULSE OXIMETER DEVICE(PROTOTYPE) TEST REPORT

Nalanda Medical College & Hospital (Covid19 ward), Patna

Doctor's name:

S.N	NAME with Designation and Address	Mobile number	Smartway Protototype Reading SPO2/BPM	Reference Pulse oximeter reading SPO2/BPM	SIGNATURE
1	Sumesh Pd. Singh Shahyahanpur. (SICU, NMCH) 78/M	9973634261	68%	63%	
2	Ajay Pd. Singh Patel Nagar, Masaulhi (SICU, NMCH) 50/M	7488898113	87%	88%	
3	Kameshwar Saw Jummar, Saharsa (SICU, NMCH) 74/M	9204730987	73%	76%	
4	Anamika Kumari Mzp, Bihar (Jyanae) 27/F	7992261070	94%	93%	
5	Rama Shankar Narpat Jang, Arania 65/M	9002074498	76%	78%	
6	Sunil Kumar Dalmiya Nagar, Rottes SICU 44/M	8252565644	88%	89%	